Equipment Purchasing

- I. The purchase of all medical equipment, furniture, fixtures and computer hardware (computers, printers, etc.) with an individual purchase price over \$5,000 or a grouping of purchases totaling over \$5,000 must be approved by the State WIC office.
- II. Purchase Approval.
 - a. Collect at least 2 price quotes from different companies.
 - b. Write a letter to the State WIC Program Manager. Include the following in the letter:
 - i. brand name of equipment
 - ii. model number
 - iii. quantity needed
 - iv. estimated cost (lower bid price or justification for higher price) supplier
 - v. justification for purchase
 - vi. where new equipment will be located
 - c. An approval letter will be sent from the State WIC Office. Keep this letter on file.
 - d. Order equipment. Add equipment information to inventory file.
- III. Equipment Criteria. The following criteria should be considered in evaluating the safety, practicality, and accuracy of equipment:

Equipment	Criteria
Infant scale - mechanical	balance beam scale with tray
	non-detachable, free-sliding weights
	 zeroing mechanism (screw type)
	can be read easily
	records to 1/2 oz or 10 gm
	can be cleaned easily
	no sharp edges on tray
Infant scale - digital	durable, sensitive, accurate and reliable
	automatically zero in between weights
	can be read easily
	• records to 1/2 oz or 10 gm
	can be cleaned easily
	no sharp edges on tray
Adult scales - mechanical	upright platform beam balance
	 non-detachable, free-sliding weights
	 zeroing mechanism (screw type)
	 records to 1/4 lb. or 0.1 kg
	can be easily read
	sturdy and functional
Adult scales - digital	 durable, sensitive, accurate and reliable
	 automatically zeroes in between weights
	• records to 1/4 lb. or 0.1 kg
	can be easily read
	sturdy and functional
Infant length measuring board	rigid headboard
	movable foot board
	measuring device marked to 1/8 inch
	no sharp edges or unfinished parts
	no loose joints
	sturdy and functional
	easily cleaned
	measurements can easily be read
Adult height measuring heard	device can be calibrated measuring device readable to 1/9 inch
Adult height measuring board	 measuring device readable to 1/8 inch measuring device perpendicular to floor
	, ,
	measuring device securely attached to walldurable and easy to clean
	 no sharp edges or unfinished parts
	 measurements can be easily read
	movable headboard at least 6 inches wide
	measuring device cannot be stretched
	device can be calibrated
	- device can be campiated